

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

APR 08 2011

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>X130305</u> <u>130305</u>	
Name <u>William Mills</u>		ID# <u>2688</u>	Date <u>4/3/11</u>
A	Agency <u>Hoonah Police Dept</u>	Phone # <u>907 945 3655</u>	
Instrument Location <u>379 Ravin Dr.</u>			
B	Alco S/N <u>X301532</u> Target Value <u>.082</u> High Pressure <u>300</u>		
Alco Test Value Average <u>.086</u> <u>.085</u>			
1 st Alco 2 nd Alco			
Signature <u>[Signature]</u>			
(OVER)			

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 15 day of JUNE, 2010. NOB

[Signature] (Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130305

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130305

APRIL 03, 2011

OPERATOR'S NAME:

MILLS WILLIAM A

OPERATOR'S NUMBER: 2688

SUBJECT'S LAST NAME:

VERIFICATION OF

SUBJECT'S FIRST NAME/NI :

CALIBRATION

D.L. #: 1234567890

DEPT/AGENCY: HNH1

CASE/REPORT: 1234567890

TEST TYPE: U

ALCO TARGET VALUE: .002

ALCO S/N: X301532

— BREATH ANALYSIS —

.002 ADJUSTED FOR 29.60 in		
ALCO TARGET	.001	16:28
BLANK TEST	.000	16:29
INTERNAL STANDARD	VERIFIED	16:29
ALCO TV 29.60 in	.006	16:30
BLANK TEST	.000	16:30
SUBJECT SAMPLE	.000	16:31
BLANK TEST	.000	16:32
ALCO TV 29.60 in	.005	16:32
BLANK TEST	.000	16:33

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130305

APRIL 03, 2011

TIME 16:34

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/00/09
HEATERS	
SAMPLE CHAMBER:	48c
BREATH TUBE:	41c
BAROMETER:	29.60 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

! " # \$ % ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ¡ ¢